


Designed by: Avelino Tintero Ferrús
Date: 15/12/2009

Approved by: Adoración Chiner Tomás
Date: 22/12/2009

Latest revision: 22/07/2015

Edition n. 3

| | |
|---|---|
| DESCRIPTION: | SPECIAL FENESTRATED SILICONE LARYNGECTOMY TUBE, N. 5 |
| MATERIAL: | NO STERILE |
| PRODUCT CLASSIFICATION | TYPE II A |
| FRANCISCO CHINER'S REFERENCE | T3051155F |
| PRODUCT CODE EAN13 | 8436530074600 |
| TECHNICAL CHARACTERISTICS OF THE PRODUCT | <p>The special fenestrated silicone laryngectomy tube, n. 5 is made up of:</p> <ul style="list-style-type: none"> • An only outer cannula: Tube and plate made of injected silicone. It is 11 mm in its outer diameter, 55 mm long and its inner diameter is 8 mm. This tube has an oval fenestration. The silicone used has the following reference: SILBIONE 71718/60/V TPTE. Medical use. |
| PACKAGING | UNIT PACKAGE |
| RECOMMENDATION | REPLACE AFTER 30 DAYS |
| EXPIRY DATE | 5 YEARS FROM DATE OF STERILIZATION, ALWAYS STORE IN ORIGINAL CONTAINER AND NORMAL CONDITIONS OF STORAGE. |
| |  |