

Designed by: Avelino Tintero Ferrús  
Date: 15/12/2009

Approved by: Adoración Chiner Tomás  
Date: 22/12/2009

Latest revision: 22/07/2015

Edition n. 3

<b>DESCRIPTION:</b>	<b>FENESTRATED SILICONE LARYNGECTOMY TUBE, N. 9</b>
<b>MATERIAL:</b>	<b>NO STERILE</b>
<b>PRODUCT CLASSIFICATION</b>	<b>TYPE II A</b>
<b>FRANCISCO CHINER'S REFERENCE</b>	<b>T3091563F</b>
<b>PRODUCT CODE EAN13</b>	<b>8470004954739</b>
<b>TECHNICAL CHARACTERISTICS OF THE PRODUCT</b>	<p>The fenestrated silicone laryngectomy tube, n. 9 is made up of:</p> <ul style="list-style-type: none"> <li>• <b>An only outer cannula:</b> Tube and plate made of injected silicone. It is 15 mm in its outer diameter, 63 mm long and its inner diameter is 12 mm. This tube has an oval fenestration. The silicone used has the following reference: <b>SILBIONE 71718/60/V TPTE. Medical use.</b></li> </ul>
<b>PACKAGING</b>	<b>UNIT PACKAGE</b>
<b>RECOMMENDATION</b>	<b>REPLACE AFTER 30 DAYS</b>
<b>EXPIRY DATE</b>	<b>5 YEARS FROM DATE OF STERILIZATION, ALWAYS STORE IN ORIGINAL CONTAINER AND NORMAL CONDITIONS OF STORAGE.</b>
	